



**Cal/EPA**

California  
Environmental  
Protection  
Agency



**Air Resources Board**

P.O. Box 2815  
2020 L Street  
Sacramento, CA  
95812-2815  
<http://www.arb.ca.gov>

November 19, 1997



Pete Wilson  
Governor

Peter M. Rooney  
Secretary for  
Environmental  
Protection

Mr. Jon Young, President  
Hasstech  
6985 Flanders Drive  
San Diego, California 92121

Dear Mr. Young:

EPA 301 Comparison Testing of VacuSmart and VacuCheck with CARB TP201.5 at  
a Facility with a Catlow ICVN System

Thank you for your patience and assistance in conducting the equivalency tests for the Hasstech (A/L) instruments. Based on the results described in the enclosed summary of the report by Hasstech's engineer, we approve the VacuCheck and VacuSmart procedures as alternatives to CARB TP-201.5, when applied to a Catlow ICVN system as described in CARB Executive Order # G-70-179.

To briefly summarize the report, comparison testing was performed using two levels of flow restriction in the front end hardware hanging at a facility with a Catlow ICVN. VacuCheck and VacuSmart passed their USEPA Method 301 comparison tests with TP201.5 for both levels of restriction.

Please contact Cindy Castronovo at (916) 263-1628 if you need further assistance.

Sincerely,

William V. Loscutoff, Chief  
Monitoring and Laboratory Division

Enclosure

cc: Jim Morgester, Chief  
Compliance Division

Jim Johnston, Chair  
CAPCOA Vapor Recovery Committee

## EPA 301 Comparison Testing of VacuSmart and VacuCheck with CARB TP201.5 at a Facility with a Catlow ICVN System

### **Introduction:**

In 1996, Hasstech requested evaluation of its VacuSmart and VacuCheck instruments as alternatives to the procedures contained in ARB TP-201.5 "Determination (by Volume Meter) of Air to Liquid Volume Ratio of Vapor Recovery Systems of Dispensing Facilities"

Section 13 of TP 201.5 "ALTERNATIVE TEST PROCEDURES" essentially states that such alternatives shall only be used with written approval from the ARB Executive Officer. The applicant is responsible for satisfying the ARB Executive Officer that the alternative certification procedure is equivalent to the subject test procedure.

ARB Testing Section staff agreed to conduct side-by-side tests of TP-201.5 to evaluate the Hasstech instruments using USEPA Method 301 as a basis for equivalency determination.

### **Test Method:**

EPA Method 301 provides four categories of procedures for determining precisions and biases which can be used to decide issues of equivalency between validated and alternative test procedures. The four general categories are:

- (1) Isotopic Spiking
- (2) Comparison Against a Validated Test Method
- (3) Analyte Spiking
- (4) Probe Placement and Arrangement for Stationary Source Stack or Duct Sampling

Only the second category applies here. Each test requires comparison runs of the validated test method and the alternative test method. Nine paired runs are required for each vapor recovery system mode.

Comparison testing was performed in two system modes using two levels of restriction to flow in the front end hardware hanging at a facility with a Catlow ICVN. This was achieved by, for instance, hanging a long narrow hose to get high restriction or a short wide hose to get low restriction.

Finally, each test mode was run for three different times; for the time it took to dispense two, three, and four gallons.

Per the rules of EPA 301, the test runs were organized into pairs of runs. Because TP-201-5 was run each time with one or more comparison runs, at least one data pair was extracted per run of TP-201-5.

Per the rules of EPA 301, the test pairs were subjected to precision and bias testing at the ninety-fifth percent confidence level with the following criteria:

(1) Precision:

- (a) If the F statistic is less than or equal to 3.44, then the precisions are adequately equal.
- (b) If the F statistic is greater than 3.44, then the candidate procedure fails.

(2) Bias:

- (a) If the t statistic is less than or equal to 1.397, then no Correction Factor is needed.
- (b) The Correction Factor is the (Validated Value) / (Candidate Value).
- (c) If a Correction Factor is at or inside the range 0.9 to 1.1, the candidate procedure passes.
- (d) If a Correction Factor is outside the range 0.9 to 1.1, the candidate procedure fails.

### Test Results

VacuSmart and VacuCheck passed their comparison tests with TP201.5 for both levels of restriction and at three refueling volumes per EPA 301 on a Catlow ICVN.

F = F statistic  
t = t statistic  
CF = Correction Factor  
VS = VacuSmart  
VC = VacuCheck

#### Low Restriction

	VS 2 gallon	VS 3 gallon	VS 4 gallon	VC 2 gallon	VC 3 gallon	VC 4 gallon
F	0.26	0.22	0.25	0.24	0.35	0.16
t	2.090	0.133	0.124	2.308	2.207	2.830
CF	0.984	1.000	1.000	1.018	1.021	1.018
	pass	pass	pass	pass	pass	pass

#### High Restriction

	VS 2 gallon	VS 3 gallon	VS 4 gallon	VC 2 gallon	VC 3 gallon	VC 4 gallon
F	0.19	0.41	0.36	3.20	3.35	3.15
t	6.054	4.864	5.205	8.392	7.399	7.410
CF	1.027	1.032	1.032	1.092	1.083	1.080
	pass	pass	pass	pass	pass	pass

#### EPA Correction Factor Averages

	VS 2 gallon	VS 3 gallon	VS 4 gallon	VC 2 gallon	VC 3 gallon	VC 4 gallon
CF	1.006	1.016	1.016	1.055	1.052	1.049
	pass	pass	pass	pass	pass	pass